



## **APPENDIX 12-2**

**NOISE METER CALIBRATION  
CERTIFICATES**

## APPENDIX 12-2. NOISE CALIBRATION CERTIFICATES



### CERTIFICATE OF CALIBRATION



**Date of Issue: 22 January 2020**

**Certificate Number: UCRT20/1094**

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way


Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Customer                    AWN Consulting  
                                   The Tecpro Building  
                                   IDA Business and Technology Park  
                                   Clonshaugh  
                                   Dublin 17

Order No.                    AWNC150120QTE  
 Description                Sound Level Meter / Pre-amp / Microphone / Associated Calibrator  
 Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00998411
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	98625
Rion	Microphone	UC-59	15917
Rion	Calibrator	NC-74	34536109
	Calibrator adaptor type if applicable		NC-74-002

Performance Class        1  
 Test Procedure            TP 2.SLM 61672-3 TPS-49  
                                   *Procedures from IEC 61672-3:2006 were used to perform the periodic tests.*  
 Type Approved to IEC 61672-1:2002    YES        Approval Number        21.21 / 13.02  
                                   *If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003*  
 Date Received            17 January 2020                            ANV Job No.        UKAS20/01036  
 Date Calibrated         22 January 2020

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

Previous Certificate	Dated	Certificate No.	Laboratory
	Initial Calibration		

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



# CERTIFICATE OF CALIBRATION



0653

**Date of Issue: 21 January 2020**

**Certificate Number: UCRT20/1093**

Issued by:

ANV Measurement Systems  
Beaufort Court  
17 Roebuck Way  
Milton Keynes MK5 8HL  
Telephone 01908 642846 Fax 01908 642814  
E-Mail: [info@noise-and-vibration.co.uk](mailto:info@noise-and-vibration.co.uk)  
Web: [www.noise-and-vibration.co.uk](http://www.noise-and-vibration.co.uk)

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Customer                    AWN Consulting  
                                  The Tecpro Building  
                                  IDA Business and Technology Park  
                                  Clonshaugh  
                                  Dublin 17

Order No.                    AWNC150120QTE  
Description                 Sound Level Meter / Pre-amp / Microphone / Associated Calibrator  
Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00998412
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	98626
Rion	Microphone	UC-59	15919
Rion	Calibrator	NC-74	34536109
	Calibrator adaptor type if applicable		NC-74-002

Performance Class        1  
Test Procedure            TP 2.SLM 61672-3 TPS-49  
*Procedures from IEC 61672-3:2006 were used to perform the periodic tests.*  
Type Approved to IEC 61672-1:2002    YES        Approval Number        21.21 / 13.02  
*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003*  
Date Received            17 January 2020                            ANV Job No.        UKAS20/01036  
Date Calibrated          21 January 2020

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

Previous Certificate	Dated	Certificate No.	Laboratory
	Initial Calibration		

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



# CERTIFICATE OF CALIBRATION



0653

**Date of Issue: 22 January 2020**

**Certificate Number: UCRT20/1096**

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer                    AWN Consulting  
                                   The Tecpro Building  
                                   IDA Business and Technology Park  
                                   Clonshaugh  
                                   Dublin 17

Order No.                    AWNC150120QTE  
 Description                Sound Level Meter / Pre-amp / Microphone / Associated Calibrator  
 Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00998410
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	98624
Rion	Microphone	UC-59	15916
Rion	Calibrator	NC-74	34536109
	Calibrator adaptor type if applicable		NC-74-002

Performance Class        1  
 Test Procedure            TP 2.SLM 61672-3 TPS-49  
                                   *Procedures from IEC 61672-3:2006 were used to perform the periodic tests.*  
 Type Approved to IEC 61672-1:2002    YES            Approval Number        21.21 / 13.02  
                                   *If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003*  
 Date Received            17 January 2020                    ANV Job No.            UKAS20/01036  
 Date Calibrated         22 January 2020

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

Previous Certificate	Dated	Certificate No.	Laboratory
	Initial Calibration		

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



# CERTIFICATE OF CALIBRATION



**Date of Issue: 15 August 2018**

**Certificate Number: UCRT18/1831**

Issued by:  
ANV Measurement Systems  
Beaufort Court  
17 Roebuck Way  
Milton Keynes MK5 8HL  
Telephone 01908 642846 Fax 01908 642814  
E-Mail: info@noise-and-vibration.co.uk  
Web: www.noise-and-vibration.co.uk  
Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages

Approved Signatory

*J. Harriman*

J. Harriman

Customer **AWN Consulting Limited**  
The Tecpro Building  
IDA Business and Technology Park  
Dublin 17  
Ireland

Order No. 1869  
Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator  
Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00586940
Rion	Firmware		1.9
Rion	Pre Amplifier	NH-25	87059
Rion	Microphone	UC-59	13402
Rion	Calibrator	NC-74	34536109
	Calibrator adaptor type if applicable		NC-74-002

Performance Class 1  
Test Procedure TP 2.SLM 61672-3 TPS-49  
*Procedures from IEC 61672-3:2006 were used to perform the periodic tests.*

Type Approved to IEC 61672-1:2002 YES Approval Number 21.21 / 13.02  
*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003*

Date Received 13 August 2018 ANV Job No. UKAS18/08513  
Date Calibrated 15 August 2018

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

Previous Certificate	Dated	Certificate No.	Laboratory
			Initial Calibration

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



# CERTIFICATE OF CALIBRATION



0653


**Date of Issue: 22 January 2020**

**Certificate Number: UCRT20/1095**

Issued by:

ANV Measurement Systems  
Beaufort Court  
17 Roebuck Way  
Milton Keynes MK5 8HL  
Telephone 01908 642846 Fax 01908 642814  
E-Mail: info@noise-and-vibration.co.uk  
Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

**Customer**                      AWN Consulting  
The Tecpro Building  
IDA Business and Technology Park  
Clonshaugh  
Dublin 17

**Order No.**                      AWNC150120QTE  
**Description**                  Sound Level Meter / Pre-amp / Microphone / Associated Calibrator  
**Identification**

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00998413
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	98627
Rion	Microphone	UC-59	15920
Rion	Calibrator	NC-74	34536109
	Calibrator adaptor type if applicable		NC-74-002

**Performance Class**        1  
**Test Procedure**            TP 2.SLM 61672-3 TPS-49  
*Procedures from IEC 61672-3:2006 were used to perform the periodic tests.*  
**Type Approved to IEC 61672-1:2002**    YES    Approval Number    21.21 / 13.02  
*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003*  
**Date Received**            17 January 2020                      ANV Job No.    UKAS20/01036  
**Date Calibrated**         22 January 2020

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

Previous Certificate	Dated	Certificate No.	Laboratory
	Initial Calibration		

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.