

Appendix 10-3

CALIBRATION CERTIFICATES

APPENDIX 10-3 CALIBRATION CERTIFICATES



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 13 January 2023

Certificate Number: UCRT23/1053

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustic Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages

Approved Signatory

K. Mistry

Customer
AWN Consulting Limited
The Teopro Building
IDA Business and Technology Park
Clonshaugh
Dublin
D17 XD90, Ireland

Order No. DOD/22/Cal045
Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator
Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00564809
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	64934
Rion	Microphone	UC-59	09447
Brüel & Kjær	Calibrator	4231	2263026
	Calibrator adaptor type if applicable		UC 0210

Performance Class

1

Test Procedure

TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013

Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received

09 January 2023

ANV Job No.

UKAS23/01009

Date Calibrated

13 January 2023

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	15 September 2020	UCRT20/1868	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION	Certificate Number UCRT23/1053
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or Internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator	Equivalent		
Customer or Lab Calibrator	Customers Calibrator		
Calibrator adaptor type if applicable	UC 0210		
Calibrator cal. date	10 January 2023		
Calibrator cert. number	UCRT23/1036		
Calibrator cal cert issued by Lab	0653		
Calibrator SPL @ STP	93.96	dB	Calibration reference sound pressure level
Calibrator frequency	999.97	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests		Start	End	
	Temperature	23.19	22.90	± 0.30 °C
	Humidity	46.5	44.5	± 3.00 %RH
	Ambient Pressure	99.66	99.71	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.0	dB	Adjusted indicated level 94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	16.9	dB A Weighting
Microphone replaced with electrical input device -		UR - Under Range Indicated	
Weighting	A	C	Z
	12.1	16.2	22.4
	dB	dB	dB
	UR	UR	UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

Prior to calibration the meter was realigned (1.0 dB drift).

..... **END**
Calibrated by: PB R 1



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 13 June 2023

Certificate Number: UCRT23/1774

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer	AWN Consulting Limited The Tecpro Building IDA Business and Technology Park Clonshaugh Dublin D17 XD90 Ireland			
Order No.	AWN200423			
Description	Sound Level Meter / Pre-amp / Microphone / Associated Calibrator			
Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00186668
	Rion	Firmware		2.1
	Rion	Pre Amplifier	NH-25	76701
	Rion	Microphone	UC-59	12813
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002
Performance Class	1			
Test Procedure	TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i>			
Type Approved to IEC 61672-1:2013	Yes <i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i>			
Date Received	13 June 2023	ANV Job No.	UKAS23/06399	
Date Calibrated	13 June 2023			

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	<i>Dated</i>	<i>Certificate No.</i>	<i>Laboratory</i>
	03 May 2022	UCRT22/1600	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION	Certificate Number UCRT23/1774
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title		NL-52/NL-42 Description for IEC 61672-1	
SLM instruction manual ref / issue		No. 56034 21-03	Source Rion
Date provided or internet download date		19 March 2021	
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator		Specified	
Customer or Lab Calibrator		Lab Calibrator	
Calibrator adaptor type if applicable		NC-74-002	
Calibrator cal. date		30 May 2023	
Calibrator cert. number		UCRT23/1727	
Calibrator cal cert issued by Lab		0653	
Calibrator SPL @ STP		94.02	dB Calibration reference sound pressure level
Calibrator frequency		1001.99	Hz Calibration check frequency
Reference level range		Single	dB
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			
Environmental conditions during tests		Start	End
Temperature		24.08	24.03
Humidity		41.0	35.8
Ambient Pressure		100.41	100.43

Indication at the Calibration Check Frequency			
Initial indicated level	94.0	dB	Adjusted indicated level 94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	19.2	dB A Weighting
Microphone replaced with electrical input device -			UR = Under Range indicated
Weighting	A	C	Z
	14.5 dB UR	16.5 dB UR	22.0 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

Prior to calibration the instrument's main PCB was replaced and the meter was realigned.

..... END
Calibrated by: K. Zablocki R 1



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 01 February 2022

Certificate Number: UCRT22/1142

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer AWN Consulting Limited
The Tecpro Building
IDA Business and Technology Park
Clonsaugh
Dublin
D17 XD90

Order No. DOD/22/Cal/038

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00998409
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	98623
	Rion	Microphone	UC-59	15915
	Brüel & Kjær	Calibrator	4231	3010472
		Calibrator adaptor type if applicable		UC 0210

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received

26 January 2022

ANV Job No.

UKAS22/01059

Date Calibrated

01 February 2022

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	07 May 2020	UCRT20/1406	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION	Certificate Number UCRT22/1142
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1
SLM instruction manual ref / issue	No. 56034 21-03 Source Rion
Date provided or internet download date	19 March 2021
	Case Corrections Wind Shield Corrections Mic Pressure to Free Field Corrections
Uncertainties provided	Yes Yes Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013	
YES	
Specified or equivalent Calibrator	Equivalent
Customer or Lab Calibrator	Customers Calibrator
Calibrator adaptor type if applicable	UC 0210
Calibrator cal. date	01 February 2022
Calibrator cert. number	UCRT22/1136
Calibrator cal cert issued by Lab	0653
Calibrator SPL @ STP	94.03 dB Calibration reference sound pressure level
Calibrator frequency	999.97 Hz Calibration check frequency
Reference level range	Single dB
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15	
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.	

Environmental conditions during tests	Start	End	
Temperature	24.53	24.56	± 0.30 °C
Humidity	40.6	39.9	± 3.00 %RH
Ambient Pressure	100.99	100.98	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.1 dB	Adjusted indicated level	94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	18.8 dB	A Weighting
Microphone replaced with electrical input device -		UR = Under Range indicated	
Weighting	A	C	Z
	11.2 dB UR	14.8 dB UR	21.3 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

..... END
Calibrated by: B. Bogdan R 3



CERTIFICATE OF CALIBRATION



Date of Issue: 04 September 2023

Certificate Number: UCRT23/2135

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer
AWN Consulting Limited
The Tecpro Building
17, Clonsaugh Business & Technology Park
Dublin

Order No. AWN160823

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00575782
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	65810
	Rion	Microphone	UC-59	19108
	Rion	Calibrator	NC-75	34724227
		Calibrator adaptor type if applicable		NC-75-022

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received 01 September 2023

ANV Job No.

UKAS23/09603

Date Calibrated 04 September 2023

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	12 July 2021	UCRT21/1841	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION	Certificate Number UCRT23/2135
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator	Specified		
Customer or Lab Calibrator	Lab Calibrator		
Calibrator adaptor type if applicable	NC-75-022		
Calibrator cal. date	04 September 2023		
Calibrator cert. number	UCRT23/2130		
Calibrator cal cert issued by Lab	0653		
Calibrator SPL @ STP	94.03	dB	Calibration reference sound pressure level
Calibrator frequency	999.99	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests	Start	End	
Temperature	24.52	24.65	± 0.30 °C
Humidity	40.8	41.8	± 3.00 %RH
Ambient Pressure	101.16	101.12	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.1	dB	Adjusted indicated level 94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	23.7	dB A Weighting
Microphone replaced with electrical input device -		UR = Under Range indicated	
Weighting	A	C	Z
	12.5	17.4	24.6
	dB	dB	dB
	UR	UR	UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

..... END
Calibrated by: K. Zablocki R 1



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 13 June 2023

Certificate Number: UCRT23/1773

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer	AWN Consulting Limited The Tecpro Building IDA Business and Technology Park Clonshaugh Dublin D17 XD90 Ireland			
Order No.	AWN200423			
Description	Sound Level Meter / Pre-amp / Microphone / Associated Calibrator			
Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00186667
	Rion	Firmware		2.1
	Rion	Pre Amplifier	NH-25	76817
	Rion	Microphone	UC-59	21140
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002
Performance Class	1			
Test Procedure	TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i>			
Type Approved to IEC 61672-1:2013	Yes	<i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i>		
Date Received	13 June 2023	ANV Job No.	UKAS23/06399	
Date Calibrated	13 June 2023			

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	<i>Dated</i>	<i>Certificate No.</i>	<i>Laboratory</i>
	12 May 2022	UCRT22/1645	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION	Certificate Number UCRT23/1773
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title NL-52/NL-42 Description for IEC 61672-1			
SLM instruction manual ref / issue No. 56034 21-03		Source Rion	
Date provided or internet download date 19 March 2021			
Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections	
Uncertainties provided Yes	Yes	Yes	
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator Specified			
Customer or Lab Calibrator Lab Calibrator			
Calibrator adaptor type if applicable NC-74-002			
Calibrator cal. date 30 May 2023			
Calibrator cert. number UCRT23/1727			
Calibrator cal cert issued by Lab 0653			
Calibrator SPL @ STP	94.02 dB	Calibration reference sound pressure level	
Calibrator frequency	1001.99 Hz	Calibration check frequency	
Reference level range	Single dB		
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			
Environmental conditions during tests		Start	End
Temperature		23.69	23.80 ± 0.30 °C
Humidity		35.1	36.7 ± 3.00 %RH
Ambient Pressure		100.41	100.42 ± 0.03 kPa
Indication at the Calibration Check Frequency			
Initial indicated level	94.0 dB	Adjusted indicated level	94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±		0.10 dB	
Self Generated Noise			
Microphone installed -	Less Than	18.4 dB	A Weighting
Microphone replaced with electrical input device -		UR = Under Range indicated	
Weighting	A	C	Z
	11.8 dB UR	15.7 dB UR	22.3 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

Prior to calibration the instrument's main PCB was replaced and the meter was realigned.

..... END
Calibrated by: C. Hirlav R 3



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 16 March 2022

Certificate Number: UCRT22/1378

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Rosbuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer	AWN Consulting Limited The Tecpro Building IDA Business and Technology Park Clonsheugh Dublin D17 XD90 Ireland			
Order No.	2201			
Description	Sound Level Meter / Pre-amp / Microphone / Associated Calibrator			
Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00998413
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	98627
	Rion	Microphone	UC-59	15920
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002
Performance Class	1			
Test Procedure	TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i>			
Type Approved to IEC 61672-1:2013	Yes <i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i>			
Date Received	16 March 2022	ANV Job No.	UKAS22/03189	
Date Calibrated	16 March 2022			

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	22 January 2020	UCRT20/1095	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION	Certificate Number UCRT22/1378
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

Sound Level Meter instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator	Specified		
Customer or Lab Calibrator	Lab Calibrator		
Calibrator adaptor type if applicable	NC-74-002		
Calibrator cal. date	17 February 2022		
Calibrator cert. number	UCRT22/1246		
Calibrator cal cert issued by Lab	0653		
Calibrator SPL @ STP	94.03	dB	Calibration reference sound pressure level
Calibrator frequency	1002.04	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests	Start	End	
Temperature	23.81	23.91	± 0.30 °C
Humidity	36.9	36.0	± 3.00 %RH
Ambient Pressure	100.50	100.50	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.3	dB	Adjusted Indicated level 94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	16.6	dB A Weighting
Microphone replaced with electrical input device -		UR = Under Range indicated	
Weighting	A	C	Z
	12.2	16.1	21.9
	dB	dB	dB
	UR	UR	UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

Calibrated by: A.Hutton

END

R 1