

Appendix 12.2 – Noise Meter Calibration Certs

Calibration Cert for Rion NL-52 Serial No. 164426



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 04 November 2021

Certificate Number: UCRT21/2362

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way


Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Customer	AWN Consulting Limited The Tecpro Building IDA Business and Technology Park Clonshaugh Dublin D17 XD90			
Order No.	DOD/21/Cal037			
Description	Sound Level Meter / Pre-amp / Microphone / Associated Calibrator			
Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00164426
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	54559
	Rion	Microphone	UC-59	09207
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002
Performance Class	1			
Test Procedure	TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i>			
Type Approved to IEC 61672-1:2013	Yes <i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i>			
Date Received	02 November 2021	ANV Job No.	UKAS21/11719	
Date Calibrated	04 November 2021			

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	<i>Dated</i>	<i>Certificate No.</i>	<i>Laboratory</i>
	05 May 2020	UCRT20/1394	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Cert for Rion NL-52 Serial No. 998413



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 16 March 2022

Certificate Number: UCRT22/1378

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way


Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Customer AWN Consulting Limited

The Tecpro Building

IDA Business and Technology Park

Clonshaugh

Dublin

D17 XD90

Ireland

Order No. 2201

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00998413
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	98627
	Rion	Microphone	UC-59	15920
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received 16 March 2022

ANV Job No. UKAS22/03189

Date Calibrated 16 March 2022

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.


Previous Certificate	Dated	Certificate No.	Laboratory
	22 January 2020	UCRT20/1095	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Cert for Rion NL-52 Serial No. 764925



CERTIFICATE OF CALIBRATION



Date of Issue: 09 September 2021

Certificate Number: UCRT21/2107

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer
AWN Consulting Limited
The Tecpro Building
IDA Business and Technology Park
Clonshaugh
Dublin
D17 XD90

Order No. 2157
Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator
Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00764925
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	65051
Rion	Microphone	UC-59	09853
Brüel & Kjær	Calibrator	4231	3010488
	Calibrator adaptor type if applicable		UC 0210

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received 03 September 2021

ANV Job No. UKAS21/09586

Date Calibrated 09 September 2021

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	10 June 2021	UCRT21/1719	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Cert for Rion NL-52 Serial No. 186669



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 12 May 2022

Certificate Number: UCRT22/1643

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer	AWN Consulting Ltd The Tecpro Building IDA Business and Technology Park Clonshaugh Dublin D17 XD90			
Order No.	DOD/22/Cal041			
Description	Sound Level Meter / Pre-amp / Microphone / Associated Calibrator			
Identification	<i>Manufacturer</i>	<i>Instrument</i>	<i>Type</i>	<i>Serial No. / Version</i>
	Rion	Sound Level Meter	NL-52	00186669
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	76819
	Rion	Microphone	UC-59	12814
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002
Performance Class	1			
Test Procedure	TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i>			
Type Approved to IEC 61672-1:2013	Yes <i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i>			
Date Received	10 May 2022	ANV Job No.	UKAS22/05320	
Date Calibrated	12 May 2022			

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	<i>Dated</i>	<i>Certificate No.</i>	<i>Laboratory</i>
	04 May 2020	UCRT20/1389	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Cert for Rion NL-52 Serial No. 998410



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 17 March 2022

Certificate Number: UCRT22/1389

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages

Approved Signatory

K. Mistry

Customer	AWN Consulting Limited The Tecpro Building IDA Business and Technology Park Clonshaugh Dublin D17 XD90 Ireland			
Order No.	2201			
Description	Sound Level Meter / Pre-amp / Microphone / Associated Calibrator			
Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00998410
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	98624
	Rion	Microphone	UC-59	15916
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002
Performance Class	1			
Test Procedure	TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i>			
Type Approved to IEC 61672-1:2013	Yes	<i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i>		
Date Received	16 March 2022	ANV Job No.	UKAS22/03189	
Date Calibrated	17 March 2022			

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	<i>Dated</i>	<i>Certificate No.</i>	<i>Laboratory</i>
	22 January 2020	UCRT20/1096	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Calibration Cert for Rion NL-52 Serial No. 164427



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 12 May 2022

Certificate Number: UCRT22/1644

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages

Approved Signatory

K. Mistry

Customer AWN Consulting Ltd
The Tecpro Building
IDA Business and Technology Park
Clonshaugh
Dublin
D17 XD90

Order No. DOD/22/Cal041

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00164427
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	54560
	Rion	Microphone	UC-59	09208
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received 10 May 2022

ANV Job No. UKAS22/05320

Date Calibrated 12 May 2022

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	05 May 2020	UCRT20/1393	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.