

**TECHNICAL APPENDIX 11.1**

**Noise Monitor In-Situ**

NML 1 – H7



NML 2 – H37



NML 3A



NML 5 – H103



NML 7 – Quarry



NML 8 – H108



NML 9 – H19



NML 831 – H14



### Noise Monitoring Locations and the Proposed Development



## TECHNICAL APPENDIX 11.2

### Wind Speed Calculations for Hub Height

A GOOD PRACTICE GUIDE TO THE APPLICATION OF ETSU-R-97 FOR THE  
ASSESSMENT AND RATING OF WIND TURBINE NOISE




#### 4.5 Wind Shear Corrections


- 4.5.1 Basing the predictions on sound power data tested in accordance with the IEC 61400-11 standard (or equivalent) should mean that the wind reference used corresponds to hub height wind speeds, standardised to 10 m height using a fixed correction (see **Annex A**). These predictions can then be compared to background levels and/or associated noise limits derived using an equivalent wind speed reference, which will have wind shear taken into account directly.
- 4.5.2 When this is not the case, for example when considering background data measured against direct wind speed measurements at 10 m height, it is necessary to apply corrections to account for this. Any such corrections should be clearly outlined and detailed in any noise assessment so that they can be reviewed by any assessor. The assessment should be made using the most detailed information available.
- 4.5.3 Examples of methods which can be used to correct predictions to account for wind shear effects, when only using a 10 m mast, are included in **Supplementary Guidance Note 4** (wind shear). This note presents methods to calculate corrections on the basis of long-term data measured at different heights, but as such data may not be available for a specific site, typical shear values are also presented. Alternatively, similarly derived corrections representing typical (average) shear values can be applied to the wind speed reference used for the derived typical background noise levels.
- 4.5.4 The following simplified method is proposed for ease of use: applying a fixed correction by subtracting the following factors from the wind speed reference used in the turbine predictions: 1 m/s for turbine hub heights of up to 30 m, 2 m/s for hub heights of up to 60 m and 3 m/s for hub heights of more than 60 m. Such a generic approach would be suitable in the context of a study made using a 10 m mast to limit costs, in the absence of site-specific data.
- 4.5.5 If it can be demonstrated that the predicted levels are below the applicable lower fixed limits regardless of wind speed, it can be seen that wind shear would not have an effect on the assessment and this may form the basis of a suitable planning condition.

**TECHNICAL APPENDIX 11.3**


**Background Noise Levels, background plus 5 trendline with the predicted noise levels against a noise limit of 45 dB(A) at each receptor**




**Chart 11.1:** NML 1 – H7, comparing the background noise level and wind speed over the measurement period




**Chart 11.2:** NML 1 – H7 for quiet daytime, background noise level, predicted level and assessment limit




**Chart 11.3:** NML 1 – H7 for night-time, background noise level, predicted level and assessment limit




**Chart 11.4:** NML 2 - H37 comparing the background noise level and wind speed over the measurement period




**Chart 11.5:** NML 2 - H37 for quiet daytime, background noise level, predicted level and assessment limit




**Chart 11.6:** NML 2 - H37 for night-time, background noise level, predicted level and assessment limit




**Chart 11.7:** NML 3A comparing the background noise level and wind speed over the measurement period




**Chart 11.8:** NML 3A for quiet daytime, background noise level, predicted level and assessment limit




**Chart 11.9:** NML 3A for night-time, background noise level, predicted level and assessment limit




**Chart 11.10:** NML 5 – H103 comparing the background noise level and wind speed over the measurement period




**Chart 11.11:** NML 5 – H103 for quiet daytime, background noise level, predicted level and assessment limit




**Chart 11.12:** NML 5 – H103 for night-time, background noise level, predicted level and assessment limit




**Chart 11.13:** NML 7 - Quarry comparing the background noise level and wind speed over the measurement period




**Chart 11.14:** NML 7 - Quarry for quiet daytime, background noise level, predicted level and assessment limit




**Chart 11.15:** NML 7 - Quarry for night-time, background noise level, predicted level and assessment limit




**Chart 11.16:** NML 8 - H108 comparing the background noise level and wind speed over the measurement period




**Chart 11.17:** NML 8 - H108 for quiet daytime, background noise level, predicted level and assessment limit




**Chart 11.18:** NML 8 - H108 for night-time, background noise level, predicted level and assessment limit




**Chart 11.19:** NML 9 - H19 comparing the background noise level and wind speed over the measurement period




**Chart 11.20:** NML 9 - H19 for quiet daytime, background noise level, predicted level and assessment limit




**Chart 11.21:** NML 9 - H19 for night-time, background noise level, predicted level and assessment limit



**Chart 11.22:** NML 831 - H14 comparing the background noise level and wind speed over the measurement period




**Chart 11.23:** NML 831 - H14 for quiet daytime, background noise level, predicted level and assessment limit




**Chart 10.24:** NML 831 - H14 for night-time, background noise level, predicted level and assessment limit



TECHNICAL APPENDIX 11.4



SoundPLAN Noise Outputs







	<p><b>MTS Calibration Ltd,</b>                  The Grange Business Centre,                  Belasis Avenue,                  Billingham TS23 1LG,                  England                  Telephone: 01642 876 410</p>																																																																				
<b>CERTIFICATE OF CALIBRATION</b>	<p>Page 1 of 11 pages</p> <p>Approved Signatory:</p> <div style="text-align: center;">                   Tony Sherris             </div>																																																																				
<p><b>Issued by:</b> MTS Calibration Ltd</p> <p><b>Date of Issue:</b> 02 August 2023      <b>Certificate Number:</b> 38648</p>																																																																					
<b>Sound Level Meter</b>																																																																					
<b>Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1</b>																																																																					
<p><b>Client:</b> Brendan Oreilly</p>	<p><b>Instrument Make:</b> Larson Davis  <b>Instrument Model:</b> LxT1L  <b>Serial Number:</b> 0005660</p>																																																																				
<b>1</b>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">Associated Equipment</th> <th style="text-align: left;">Make</th> <th style="text-align: left;">Model</th> <th style="text-align: left;">Serial number</th> </tr> </thead> <tbody> <tr> <td>Pre-amplifier</td> <td>Larson Davis</td> <td>PRMLxT1L</td> <td>055684</td> </tr> <tr> <td>Microphone</td> <td>PCB</td> <td>377B02</td> <td>305875</td> </tr> <tr> <td>Calibrator</td> <td>Larson Davis</td> <td>CAL200</td> <td>9175</td> </tr> <tr> <td>Calibrator supplied by</td> <td colspan="3">MTS for this calibration</td> </tr> </tbody> </table>	Associated Equipment	Make	Model	Serial number	Pre-amplifier	Larson Davis	PRMLxT1L	055684	Microphone	PCB	377B02	305875	Calibrator	Larson Davis	CAL200	9175	Calibrator supplied by	MTS for this calibration																																																		
Associated Equipment	Make	Model	Serial number																																																																		
Pre-amplifier	Larson Davis	PRMLxT1L	055684																																																																		
Microphone	PCB	377B02	305875																																																																		
Calibrator	Larson Davis	CAL200	9175																																																																		
Calibrator supplied by	MTS for this calibration																																																																				
The measurements were performed at The Grange Business Centre, Belasis Avenue, TS23 1LD. The results only apply to the items tested.																																																																					
Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1																																																																					
Test results summary, detailed results are shown on subsequent pages.																																																																					
<b>Tests performed</b>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">Section</th> <th style="text-align: left;">Results of test</th> <th style="text-align: left;">Page</th> <th style="text-align: left;">Comments</th> </tr> </thead> <tbody> <tr> <td>Calibration Certificate</td> <td></td> <td>1</td> <td></td> </tr> <tr> <td>Additional information</td> <td></td> <td>2</td> <td></td> </tr> <tr> <td>Indication with Calibrator Supplied</td> <td>No Limit</td> <td>3</td> <td></td> </tr> <tr> <td>Self-Generated Noise</td> <td>No Limit</td> <td>3</td> <td></td> </tr> <tr> <td>Frequency and Time-weightings at 1kHz</td> <td>Complies</td> <td>3</td> <td></td> </tr> <tr> <td>Long term stability</td> <td>Complies</td> <td>3</td> <td></td> </tr> <tr> <td>High stability</td> <td>Complies</td> <td>3</td> <td></td> </tr> <tr> <td>Acoustic Tests</td> <td>Complies</td> <td>4</td> <td></td> </tr> <tr> <td>Frequency Weighting A</td> <td>Complies</td> <td>5</td> <td></td> </tr> <tr> <td>Frequency Weighting C</td> <td>Complies</td> <td>6</td> <td></td> </tr> <tr> <td>Frequency Weighting Z</td> <td>Complies</td> <td>7</td> <td></td> </tr> <tr> <td>Level Linearity</td> <td>Complies</td> <td>8</td> <td></td> </tr> <tr> <td>Level Linearity Range Control</td> <td>n/a</td> <td></td> <td>SLM only has one range</td> </tr> <tr> <td>Tone-burst Response</td> <td>Complies</td> <td>9</td> <td></td> </tr> <tr> <td>Peak C sound level</td> <td>Complies</td> <td>10</td> <td></td> </tr> <tr> <td>Overload indication</td> <td>Complies</td> <td>11</td> <td></td> </tr> </tbody> </table>	Section	Results of test	Page	Comments	Calibration Certificate		1		Additional information		2		Indication with Calibrator Supplied	No Limit	3		Self-Generated Noise	No Limit	3		Frequency and Time-weightings at 1kHz	Complies	3		Long term stability	Complies	3		High stability	Complies	3		Acoustic Tests	Complies	4		Frequency Weighting A	Complies	5		Frequency Weighting C	Complies	6		Frequency Weighting Z	Complies	7		Level Linearity	Complies	8		Level Linearity Range Control	n/a		SLM only has one range	Tone-burst Response	Complies	9		Peak C sound level	Complies	10		Overload indication	Complies	11	
Section	Results of test	Page	Comments																																																																		
Calibration Certificate		1																																																																			
Additional information		2																																																																			
Indication with Calibrator Supplied	No Limit	3																																																																			
Self-Generated Noise	No Limit	3																																																																			
Frequency and Time-weightings at 1kHz	Complies	3																																																																			
Long term stability	Complies	3																																																																			
High stability	Complies	3																																																																			
Acoustic Tests	Complies	4																																																																			
Frequency Weighting A	Complies	5																																																																			
Frequency Weighting C	Complies	6																																																																			
Frequency Weighting Z	Complies	7																																																																			
Level Linearity	Complies	8																																																																			
Level Linearity Range Control	n/a		SLM only has one range																																																																		
Tone-burst Response	Complies	9																																																																			
Peak C sound level	Complies	10																																																																			
Overload indication	Complies	11																																																																			
The instrument was within the above specification as received - no modifications were made																																																																					
The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013																																																																					
<p><b>Additional tests performed</b></p> <p>Microphone full frequency response</p> <p>Filter calibration, third octave or octave</p>	<table style="width: 100%;"> <tr> <td style="text-align: center;"><b>Reference</b></td> <td></td> </tr> <tr> <td style="text-align: center;">38650</td> <td style="text-align: right;">See additional certificate</td> </tr> <tr> <td style="text-align: center;">38648F</td> <td style="text-align: right;">See additional certificate</td> </tr> </table>	<b>Reference</b>		38650	See additional certificate	38648F	See additional certificate																																																														
<b>Reference</b>																																																																					
38650	See additional certificate																																																																				
38648F	See additional certificate																																																																				
This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.																																																																					

	<b>MTS Calibration Ltd,</b> The Grange Business Centre, Belasis Avenue, Billingham TS23 1LG, England Telephone: 01642 876 410																																																																				
<b>CERTIFICATE OF CALIBRATION</b>	Page 1 of 11 pages Approved Signatory:  Tony Sherris																																																																				
Issued by: <b>MTS Calibration Ltd</b>  Date of Issue: <b>07 March 2022</b> Certificate Number: <b>38175</b>																																																																					
<b>Sound Level Meter</b>																																																																					
<b>Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1</b>																																																																					
Client: <b>Brendan Oreilly</b>	Instrument Make: <b>Larson Davis</b> Instrument Model: <b>LxT1L</b> Serial Number: <b>0005058</b>																																																																				
<b>2</b>	<table border="0" style="width: 100%;"> <thead> <tr> <th style="text-align: left;">Associated Equipment</th> <th style="text-align: left;">Make</th> <th style="text-align: left;">Model</th> <th style="text-align: left;">Serial number</th> </tr> </thead> <tbody> <tr> <td>Preamplifier</td> <td>PCB</td> <td>PRMLxT1L</td> <td>036054</td> </tr> <tr> <td>Microphone</td> <td>PCB</td> <td>377B02</td> <td>153775</td> </tr> <tr> <td>Calibrator</td> <td>Larson Davis</td> <td>CAL200</td> <td>9175</td> </tr> <tr> <td>Calibrator supplied by</td> <td colspan="3">MTS for this calibration</td> </tr> </tbody> </table>	Associated Equipment	Make	Model	Serial number	Preamplifier	PCB	PRMLxT1L	036054	Microphone	PCB	377B02	153775	Calibrator	Larson Davis	CAL200	9175	Calibrator supplied by	MTS for this calibration																																																		
Associated Equipment	Make	Model	Serial number																																																																		
Preamplifier	PCB	PRMLxT1L	036054																																																																		
Microphone	PCB	377B02	153775																																																																		
Calibrator	Larson Davis	CAL200	9175																																																																		
Calibrator supplied by	MTS for this calibration																																																																				
The measurements were performed at The Grange Business Centre, Belasis Avenue, TS23 1LD. The results only apply to the items tested.																																																																					
<b>Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1</b>																																																																					
<b>Test results summary, detailed results are shown on subsequent pages.</b>																																																																					
<b>Tests performed</b>	<table border="0" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">Section</th> <th style="text-align: left;">Results of test</th> <th style="text-align: left;">Page</th> <th style="text-align: left;">Comments</th> </tr> </thead> <tbody> <tr> <td>Calibration Certificate</td> <td></td> <td>1</td> <td></td> </tr> <tr> <td>Additional information</td> <td></td> <td>2</td> <td></td> </tr> <tr> <td>Indication with Calibrator Supplied</td> <td><b>No Limit</b></td> <td>3</td> <td></td> </tr> <tr> <td>Self-Generated Noise</td> <td><b>No Limit</b></td> <td>3</td> <td></td> </tr> <tr> <td>Frequency and Time-weightings at 1kHz</td> <td><b>Complies</b></td> <td>3</td> <td></td> </tr> <tr> <td>Long term stability</td> <td><b>Complies</b></td> <td>3</td> <td></td> </tr> <tr> <td>High stability</td> <td><b>Complies</b></td> <td>3</td> <td></td> </tr> <tr> <td>Acoustic Tests</td> <td><b>Complies</b></td> <td>4</td> <td></td> </tr> <tr> <td>Frequency Weighting A</td> <td><b>Complies</b></td> <td>5</td> <td></td> </tr> <tr> <td>Frequency Weighting C</td> <td><b>Complies</b></td> <td>6</td> <td></td> </tr> <tr> <td>Frequency Weighting Z</td> <td><b>Complies</b></td> <td>7</td> <td></td> </tr> <tr> <td>Level Linearity</td> <td><b>Complies</b></td> <td>8</td> <td></td> </tr> <tr> <td>Level Linearity Range Control</td> <td></td> <td>n/a</td> <td>SLM only has one range</td> </tr> <tr> <td>Tone-burst Response</td> <td><b>Complies</b></td> <td>9</td> <td></td> </tr> <tr> <td>Peak C sound level</td> <td><b>Complies</b></td> <td>10</td> <td></td> </tr> <tr> <td>Overload indication</td> <td><b>Complies</b></td> <td>11</td> <td></td> </tr> </tbody> </table>	Section	Results of test	Page	Comments	Calibration Certificate		1		Additional information		2		Indication with Calibrator Supplied	<b>No Limit</b>	3		Self-Generated Noise	<b>No Limit</b>	3		Frequency and Time-weightings at 1kHz	<b>Complies</b>	3		Long term stability	<b>Complies</b>	3		High stability	<b>Complies</b>	3		Acoustic Tests	<b>Complies</b>	4		Frequency Weighting A	<b>Complies</b>	5		Frequency Weighting C	<b>Complies</b>	6		Frequency Weighting Z	<b>Complies</b>	7		Level Linearity	<b>Complies</b>	8		Level Linearity Range Control		n/a	SLM only has one range	Tone-burst Response	<b>Complies</b>	9		Peak C sound level	<b>Complies</b>	10		Overload indication	<b>Complies</b>	11	
Section	Results of test	Page	Comments																																																																		
Calibration Certificate		1																																																																			
Additional information		2																																																																			
Indication with Calibrator Supplied	<b>No Limit</b>	3																																																																			
Self-Generated Noise	<b>No Limit</b>	3																																																																			
Frequency and Time-weightings at 1kHz	<b>Complies</b>	3																																																																			
Long term stability	<b>Complies</b>	3																																																																			
High stability	<b>Complies</b>	3																																																																			
Acoustic Tests	<b>Complies</b>	4																																																																			
Frequency Weighting A	<b>Complies</b>	5																																																																			
Frequency Weighting C	<b>Complies</b>	6																																																																			
Frequency Weighting Z	<b>Complies</b>	7																																																																			
Level Linearity	<b>Complies</b>	8																																																																			
Level Linearity Range Control		n/a	SLM only has one range																																																																		
Tone-burst Response	<b>Complies</b>	9																																																																			
Peak C sound level	<b>Complies</b>	10																																																																			
Overload indication	<b>Complies</b>	11																																																																			
The instrument was within the above specification as received - no modifications were made																																																																					
The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013																																																																					
<b>Additional tests performed</b>	<table border="0" style="width: 100%;"> <thead> <tr> <th style="text-align: left;">Reference</th> <th style="text-align: left;"></th> </tr> </thead> <tbody> <tr> <td>38177</td> <td>See additional certificate</td> </tr> <tr> <td>38175F</td> <td>See additional certificate</td> </tr> </tbody> </table>	Reference		38177	See additional certificate	38175F	See additional certificate																																																														
Reference																																																																					
38177	See additional certificate																																																																				
38175F	See additional certificate																																																																				
This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.																																																																					



**MTS Calibration Ltd,**  
**The Grange Business Centre,**  
**Belasis Avenue,**  
**Billingham TS23 1LG,**  
**England**  
**Telephone: 01642 876 410**

**CERTIFICATE OF CALIBRATION**

Page 1 of 11 pages

Issued by: **MTS Calibration Ltd**

Approved Signatory:

Date of Issue: **05 August 2022**

Certificate Number: **37266**

**Tony Sherris**

**Sound Level Meter**

**Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1**

**Client:** Environmental Measurements  
 Unit 12, Tallaght Business Centre  
 Whitestown Business Park  
 Co.Dublin 24, Ireland

**Instrument Make:** Larson Davis  
**Instrument Model:** LxT1L  
**Serial Number:** 0005990

**4**

Associated Equipment	Make	Model	Serial number
Preamplifier	Larson Davis	PRMLxT1L	055804
Microphone	PCB	377B02	316349
Calibrator	Brüel & Kjær	4231	3014620
Calibrator supplied by	MTS for this calibration		

The measurements were performed at Elvington Close, Billingham, TS23 3YS. The results only apply to the items tested.

**Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1**

Test results summary, detailed results are shown on subsequent pages.


Tests performed	Section	Results of test	Page	Comments
Calibration Certificate	22		1	
Additional information			2	
Indication with Calibrator Supplied	10	No Limit	3	
Self-Generated Noise	11	No Limit	3	
Frequency and Time-weightings at 1kHz	14	Complies	3	
Long term stability	15	Complies	3	
High stability	21	Complies	3	
Acoustic Tests	12	Complies	4	
Frequency Weighting A	13	Complies	5	
Frequency Weighting C	13	Complies	6	
Frequency Weighting Z	13	Complies	7	
Level Linearity	16	Complies	8	
Level Linearity Range Control	17	n/a		SLM only has one range
Tone-burst Response	18	Complies	9	
Peak C sound level	19	Complies	10	
Overload indication	20	Complies	11	

The instrument was within the above specification as received - no modifications were made

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013

Additional tests performed	Reference	
Microphone full frequency response	37268	See additional certificate
Filter calibration, third octave or octave	37266F	See additional certificate

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.




**MTS Calibration Ltd,**  
**The Grange Business Centre,**  
**Belasis Avenue,**  
**Billingham TS23 1LG,**  
**England**  
**Telephone: 01642 876 410**

---

## CERTIFICATE OF CALIBRATION

**Page 1 of 11 pages**

Approved Signatory:

  
 Tony Sherris

**Issued by:** MTS Calibration Ltd

**Date of Issue:** 08 March 2023

**Certificate Number:** 38149

### Sound Level Meter

#### Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1

**Client:** Brendan O'Reilly

**Instrument Make:** Larson Davis  
**Instrument Model:** LxT1L  
**Serial Number:** 0005992

	Associated Equipment	Make	Model	Serial number
<b>6</b>	Preamplifier	PCB	PRMLxT1L	028029
	Microphone	PCB	377B02	147913
	Calibrator	Larson Davis	CAL200	18140
	Calibrator supplied by the Client, with the SLM			

The measurements were performed at The Grange Business Centre, Belasis Avenue, TS23 1LD. The results only apply to the items tested.

Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1

Test results summary, detailed results are shown on subsequent pages.

Tests performed	Section	Results of test	Page	Comments
Calibration Certificate	22		1	
Additional information			2	
Indication with Calibrator Supplied	10	No Limit	3	
Self-Generated Noise	11	No Limit	3	
Frequency and Time-weightings at 1kHz	14	Complies	3	
Long term stability	15	Complies	3	
High stability	21	Complies	3	
Acoustic Tests	12	Complies	4	
Frequency Weighting A	13	Complies	5	
Frequency Weighting C	13	Complies	6	
Frequency Weighting Z	13	Complies	7	
Level Linearity	16	Complies	8	
Level Linearity Range Control	17	n/a		SLM only has one range
Tone-burst Response	18	Complies	9	
Peak C sound level	19	Complies	10	
Overload indication	20	Complies	11	

The instrument was within the above specification as received - no modifications were made

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013

Additional tests performed	Reference	
Microphone full frequency response	38151	See additional certificate
Filter calibration, third octave or octave	38149F	See additional certificate
Calibrator calibration	38152U	See additional UKAS certificate

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



**MTS Calibration Ltd,**  
**The Grange Business Centre,**  
**Belasis Avenue,**  
**Billingham TS23 1LG,**  
**England**  
**Telephone: 01642 876 410**

**CERTIFICATE OF CALIBRATION**

Page 1 of 11 pages

Issued by: **MTS Calibration Ltd**

Approved Signatory:

Date of Issue: **05 August 2022** Certificate Number: **37269**

**Tony Sherris**

**Sound Level Meter**

**Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1**

**Client:** Environmental Measurements for Brendan O'Reilly  
 Unit 12, Tallaght Business Centre  
 Whitestown Business Park  
 Co.Dublin 24, Ireland

**Instrument Make:** Larson Davis  
**Instrument Model:** LxT1L  
**Serial Number:** 0005046

**7**

Associated Equipment	Make	Model	Serial number
Preamplifier	Larson Davis	PRMLxT1L	042734
Microphone	PCB	377B02	172753
Calibrator	Brüel & Kjær	4231	2343058
Calibrator supplied by	MTS for this calibration		

The measurements were performed at Elvington Close, Billingham, TS23 3YS. The results only apply to the items tested.

**Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1**

Test results summary, detailed results are shown on subsequent pages.

Tests performed	Section	Results of test	Page	Comments
Calibration Certificate	22		1	
Additional information			2	
Indication with Calibrator Supplied	10	No Limit	3	
Self-Generated Noise	11	No Limit	3	
Frequency and Time-weightings at 1kHz	14	Complies	3	
Long term stability	15	Complies	3	
High stability	21	Complies	3	
Acoustic Tests	12	Complies	4	
Frequency Weighting A	13	Complies	5	
Frequency Weighting C	13	Complies	6	
Frequency Weighting Z	13	Complies	7	
Level Linearity	16	Complies	8	
Level Linearity Range Control	17	n/a		SLM only has one range
Tone-burst Response	18	Complies	9	
Peak C sound level	19	Complies	10	
Overload indication	20	Complies	11	

The instrument was within the above specification as received - no modifications were made

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013

Additional tests performed	Reference	
Microphone full frequency response	37271	See additional certificate
Filter calibration, third octave or octave	37269F	See additional certificate

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# Calibration Certificate

**Certificate Number** 2021014389

**Customer:**

Environmental Measurement  
Unit 12 Tallaght Business Centre  
Whitestown Business Park  
Dublin, 24, Ireland

**Model Number** LxT SE  
**Serial Number** 0006871  
**Test Results** **Pass**

**Initial Condition** As Manufactured

**Description** Sound Expert LxT  
Class 1 Sound Level Meter  
Firmware Revision: 2,404

**Procedure Number** D0001.8378  
**Technician** Ron Harris  
**Calibration Date** 10 Nov 2021

**Calibration Due**  
**Temperature** 23.46 °C ± 0.25 °C  
**Humidity** 51.9 %RH ± 2.0 %RH  
**Static Pressure** 86.64 kPa ± 0.13 kPa

**Evaluation Method** Tested electrically using Larson Davis PRMLxT1L S/N 070117 and a 12.0 pF capacitor to simulate microphone capacitance. Data reported in dB re 20 µPa assuming a microphone sensitivity of 23.6 mV/Pa.

**Compliance Standards** Compliant to Manufacturer Specifications and the following standards when combined with Calibration Certificate from procedure D0001.8384:

IEC 60651:2001 Type 1	ANSI S1.4-2014 Class 1
IEC 60804:2000 Type 1	ANSI S1.4 (R2006) Type 1
IEC 61252:2002	ANSI S1.25 (R2007)
IEC 61672:2013 Class 1	ANSI S1.43 (R2007) Type 1
IEC 61260:2001 Class 1	ANSI S1.11 (R2009) Class 1

Issuing lab certifies that the instrument described above meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST), or other national measurement institutes, and meets the requirements of ISO/IEC 17025:2017. **Test points marked with a ‡ in the uncertainties column do not fall within this laboratory's scope of accreditation.**

The quality system is registered to ISO 9001:2015.

This calibration is a direct comparison of the unit under test to the listed reference standards and did not involve any sampling plans to complete. No allowance has been made for the instability of the test device due to use, time, etc. Such allowances would be made by the customer as needed.

The uncertainties were computed in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). A coverage factor of approximately 2 sigma (k=2) has been applied to the standard uncertainty to express the expanded uncertainty at approximately 95% confidence level.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the organization issuing this report.

Correction data from Larson Davis LxT Manual for SoundTrack LxT & SoundExpert Lxt, I770.01 Rev O Supporting Firmware Version 4.0.5, 2019-09-10

Calibration Check Frequency: 1000 Hz; Reference Sound Pressure Level: 114 dB re 20 µPa

LARSON DAVIS – A PCB PIEZOTRONICS DIV.  
1681 West 820 North  
Provo, UT 84601, United States  
716-684-0001



2021-11-10T09:02:03

Page 1 of 8

D0001.8407 Rev F

# Calibration Certificate

**Certificate Number** 2021014391

**Customer:**

Environmental Measurement  
Unit 12 Tallaght Business Centre  
Whitestown Business Park  
Dublin, 24, Ireland

<b>Model Number</b>	LxT SE	<b>Procedure Number</b>	D0001.8378
<b>Serial Number</b>	0006869	<b>Technician</b>	Ron Harris
<b>Test Results</b>	<b>Pass</b>	<b>Calibration Date</b>	10 Nov 2021
<b>Initial Condition</b>	As Manufactured	<b>Calibration Due</b>	
<b>Description</b>	Sound Expert LxT Class 1 Sound Level Meter Firmware Revision: 2,404	<b>Temperature</b>	23.41 °C ± 0.25 °C
		<b>Humidity</b>	53.2 %RH ± 2.0 %RH
		<b>Static Pressure</b>	86.63 kPa ± 0.13 kPa

**Evaluation Method** Tested electrically using Larson Davis PRMLxT1L S/N 070113 and a 12.0 pF capacitor to simulate microphone capacitance. Data reported in dB re 20 µPa assuming a microphone sensitivity of 23.6 mV/Pa.

**Compliance Standards** Compliant to Manufacturer Specifications and the following standards when combined with Calibration Certificate from procedure D0001,8384:

IEC 60651:2001 Type 1	ANSI S1.4-2014 Class 1
IEC 60804:2000 Type 1	ANSI S1.4 (R2006) Type 1
IEC 61252:2002	ANSI S1.25 (R2007)
IEC 61672:2013 Class 1	ANSI S1.43 (R2007) Type 1
IEC 61260:2001 Class 1	ANSI S1.11 (R2009) Class 1

Issuing lab certifies that the instrument described above meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST), or other national measurement institutes, and meets the requirements of ISO/IEC 17025:2017. **Test points marked with a ‡ in the uncertainties column do not fall within this laboratory's scope of accreditation.**

The quality system is registered to ISO 9001:2015.

This calibration is a direct comparison of the unit under test to the listed reference standards and did not involve any sampling plans to complete. No allowance has been made for the instability of the test device due to use, time, etc. Such allowances would be made by the customer as needed.

The uncertainties were computed in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). A coverage factor of approximately 2 sigma (k=2) has been applied to the standard uncertainty to express the expanded uncertainty at approximately 95% confidence level.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the organization issuing this report.

Correction data from Larson Davis LxT Manual for SoundTrack LxT & SoundExpert Lxt, I770.01 Rev O Supporting Firmware Version 4.0.5, 2019-09-10

Calibration Check Frequency: 1000 Hz; Reference Sound Pressure Level: 114 dB re 20 µPa

LARSON DAVIS - A PCB PIEZOTRONICS DIV.

1681 West 820 North  
Provo, UT 84601, United States  
716-684-0001

2021-11-10T09:15:37



# Calibration Certificate

**Certificate Number** 2021014392

**Customer:**

Environmental Measurement  
Unit 12 Tallaght Business Centre  
Whitestown Business Park  
Dublin, 24, Ireland

**Model Number** LxT SE  
**Serial Number** 0006870  
**Test Results** **Pass**

**Initial Condition** As Manufactured

**Description** Sound Expert LxT  
Class 1 Sound Level Meter  
Firmware Revision: 2,404

**Procedure Number** D0001.8378  
**Technician** Ron Harris  
**Calibration Date** 10 Nov 2021  
**Calibration Due**  
**Temperature** 23,5 °C ± 0,25 °C  
**Humidity** 53 %RH ± 2,0 %RH  
**Static Pressure** 86.69 kPa ± 0.13 kPa

**Evaluation Method** Tested electrically using Larson Davis PRMLxT1L S/N 070116 and a 12.0 pF capacitor to simulate microphone capacitance. Data reported in dB re 20 µPa assuming a microphone sensitivity of 23.6 mV/Pa.

**Compliance Standards** Compliant to Manufacturer Specifications and the following standards when combined with Calibration Certificate from procedure D0001.8384:

IEC 60651:2001 Type 1	ANSI S1.4-2014 Class 1
IEC 60804:2000 Type 1	ANSI S1.4 (R2006) Type 1
IEC 61252:2002	ANSI S1.25 (R2007)
IEC 61672:2013 Class 1	ANSI S1.43 (R2007) Type 1
IEC 61260:2001 Class 1	ANSI S1.11 (R2009) Class 1

Issuing lab certifies that the instrument described above meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST), or other national measurement institutes, and meets the requirements of ISO/IEC 17025:2017. **Test points marked with a ‡ in the uncertainties column do not fall within this laboratory's scope of accreditation.**

The quality system is registered to ISO 9001:2015.

This calibration is a direct comparison of the unit under test to the listed reference standards and did not involve any sampling plans to complete. No allowance has been made for the instability of the test device due to use, time, etc. Such allowances would be made by the customer as needed.

The uncertainties were computed in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). A coverage factor of approximately 2 sigma (k=2) has been applied to the standard uncertainty to express the expanded uncertainty at approximately 95% confidence level.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the organization issuing this report.

Correction data from Larson Davis LxT Manual for SoundTrack LxT & SoundExpert Lxt, I770.01 Rev O Supporting Firmware Version 4,0,5, 2019-09-10

Calibration Check Frequency: 1000 Hz; Reference Sound Pressure Level: 114 dB re 20 µPa

LARSON DAVIS – A PCB PIEZOTRONICS DIV.  
1681 West 820 North  
Provo, UT 84601, United States  
716-684-0001

2021-11-10T09:41:16



**TECHNICAL APPENDIX 11.6**

**Candidate Turbine Manufacturer's Performance Specification**

**RESTRICTED**

Document no.: 0090-2475 V02

Document owner: Platform Management

Type: T05 - General Description

Performance Specification V117-4.3 MW 50/60 Hz  
Power Curves, Ct Values and Sound Curves for Power  
Optimized (PO) Mode

Date: 2020-01-03

Restricted

Page 11 of 19

**4.3 Sound Curves, Power Optimized Mode PO2/PO2-0S;  
HWO Disabled**

Sound Power Level at Hub Height		
Conditions for Sound Power Level:	Measurement standard IEC 61400-11 ed. 3 Maximum turbulence at hub height: 30% Inflow angle (vertical): 0 ±2° Air density: 1.225 kg/m³	
Wind speed at hub height [m/s]	Sound Power Level at Hub Height [dBA] Mode PO2 (Blades with serrated trailing edge)	Sound Power Level at Hub Height [dBA] Mode PO2-0S (Blades without serrated trailing edge)
3	91.7	93.7
4	92.2	94.6
5	94	97.1
6	97	100.2
7	100	103.1
8	102.8	105.6
9	105.1	107.7
10	106	108.5
11	106	108.5
12	106	108.5
13	106	108.5
14	106	108.5
15	106	108.5
16	106	108.5
17	106	108.5
18	106	108.5
19	106	108.5
20	106	108.5
21	106	108.5
22	106	108.5
23	106	108.5
24	106	108.5
25	106	108.5
26	106	108.5
27	106	108.5

Table 4-3: Sound curves, Mode PO2/PO2-0S (High Wind Operation Disabled)

Vestas Wind Systems A/S · Hedeager 42 · 8200 Aarhus N · Denmark · www.vestas.com

Classification: Restricted



**VESTAS PROPRIETARY NOTICE**

Original Instruction: T05 0090-2475 VER 02

T05 0090-2475 Ver 02 - Approved- Exported from DMS: 2024-01-29 by FIGRL

## Candidate Turbine Manufacturer's Performance Specifications during Curtailment Strategy

### 10.3 Sound Curves, Sound Optimized Mode SO1

Sound Power Level at Hub Height	
Conditions for Sound Power Level:	Measurement standard IEC 61400-11 ed. 3 Maximum turbulence at hub height: 30% Inflow angle (vertical): $0 \pm 2^\circ$ Air density: $1.225 \text{ kg/m}^3$
Wind speed at hub height [m/s]	Sound Power Level at Hub Height [dBA] Sound Optimized Mode SO1 (Blades with serrated trailing edge)
3	92.2
4	92.8
5	94.0
6	97.0
7	100.0
8	102.7
9	104.2
10	105.0
11	105.0
12	105.0
13	105.0
14	105.0
15	105.0
16	105.0
17	105.0
18	105.0
19	105.0
20	105.0

Table 10-3: Sound curves, Sound Optimized Mode SO1

### 12.3 Sound Curves, Sound Optimized Mode SO2

Sound Power Level at Hub Height	
Conditions for Sound Power Level:	Measurement standard IEC 61400-11 ed. 3 Maximum turbulence at hub height: 30% Inflow angle (vertical): $0 \pm 2^\circ$ Air density: 1.225 kg/m <sup>3</sup>
Wind speed at hub height [m/s]	Sound Power Level at Hub Height [dBA] Sound Optimized Mode SO2 (Blades with serrated trailing edge)
3	92.2
4	92.8
5	94.0
6	97.0
7	99.9
8	101.6
9	102.3
10	102.3
11	102.4
12	102.7
13	103.0
14	103.0
15	103.0
16	103.0
17	103.0
18	103.0
19	103.0
20	103.0

Table 12-3: Sound curves, Sound Optimized Mode SO2