



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 20 April 2018

Certificate Number: UCRT18/1432

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way


Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

| |
|-------------------------------------------------------------------------------------|
| Page 1 of 2 Pages |
| Approved Signatory |
|  |
| K. Mistry |

Customer **AWN Consulting Limited**
 The Tecpro Building
 Clonsaugh Business and Technology Park
 Dublin 17
 Ireland
 D17 NX50

Order No.

Description **Sound Level Meter / Pre-amp / Microphone / Associated Calibrator**
Identification

| <i>Manufacturer</i> | <i>Instrument</i> | <i>Type</i> | <i>Serial No. / Version</i> |
|---------------------|---------------------------------------|-------------|-----------------------------|
| Rion | Sound Level Meter | NL-52 | 00186667 |
| Rion | Firmware | | 1.9 |
| Rion | Pre Amplifier | NH-25 | 76817 |
| Rion | Microphone | UC-59 | 12812 |
| Rion | Calibrator | NC-74 | 34536109 |
| | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49
Procedures from IEC 61672-3:2006 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2002 YES Approval Number 21.21 / 13.02
If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 20 April 2018 ANV Job No. UKAS18/04261

Date Calibrated 20 April 2018

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

| Previous Certificate | Dated | Certificate No. | Laboratory |
|----------------------|---------------------|-----------------|------------|
| | Initial Calibration | | |

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number
UCRT18/1432

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | |
|--------------------------------------------------------------------------|-------------------|-----------------------------------------------|
| SLM instruction manual title | Sound Level Meter | NL-42 / NL-52 |
| SLM instruction manual ref / issue | | 11-03 |
| SLM instruction manual source | Manufacturer | |
| Internet download date if applicable | N/A | |
| Case corrections available | Yes | |
| Uncertainties of case corrections | Yes | |
| Source of case data | Manufacturer | |
| Wind screen corrections available | Yes | |
| Uncertainties of wind screen corrections | Yes | |
| Source of wind screen data | Manufacturer | |
| Mic pressure to free field corrections | Yes | |
| Uncertainties of Mic to F.F. corrections | Yes | |
| Source of Mic to F.F. corrections | Manufacturer | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | | Yes |
| Specified or equivalent Calibrator | Specified | |
| Customer or Lab Calibrator | Lab Calibrator | |
| Calibrator adaptor type if applicable | NC-74-002 | |
| Calibrator cal. date | 05 April 2018 | |
| Calibrator cert. number | UCRT18/1348 | |
| Calibrator cal cert issued by | 0653 | |
| Calibrator SPL @ STP | 93.98 | dB Calibration reference sound pressure level |
| Calibrator frequency | 1001.90 | Hz Calibration check frequency |
| Reference level range | 25 - 130 | dB |

Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15
 Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 22.18 | 22.93 | ± 0.30 °C |
| Humidity | 53.1 | 51.5 | ± 3.00 %RH |
| Ambient Pressure | 101.32 | 101.34 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|------------------------------------------------------------------------------------|------|----|--------------------------|------|----|
| Initial indicated level | 94.0 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|----------------------------------------------------------------|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

| Microphone replaced with electrical input device - | | UR = Under Range indicated | | | | | | | |
|------------------------------------------------------|------|----------------------------|----|------|----|----|------|----|----|
| Weighting | A | dB | UR | C | dB | UR | Z | dB | UR |
| | 12.7 | | | 16.7 | | | 22.7 | | |
| Uncertainty of the electrical self generated noise ± | | 0.12 dB | | | | | | | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

..... END

Calibrated by: A Patel

R 1

Additional Comments

None



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 20 April 2018

Certificate Number: UCRT18/1440

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

| |
|--------------------|
| Page 1 of 2 Pages |
| Approved Signatory |
| |
| K. Mistry |

Customer AWN Consulting Limited
 The Tecpro Building
 Clonshaugh Business and Technology Park
 Dublin 17
 Ireland
 D17 NX50

Order No.

| Description | Sound Level Meter / Pre-amp / Microphone / Associated Calibrator | | | |
|----------------|------------------------------------------------------------------|---------------------------------------|-------|----------------------|
| Identification | Manufacturer | Instrument | Type | Serial No. / Version |
| | Rion | Sound Level Meter | NL-52 | 00186669 |
| | Rion | Firmware | | 1.9 |
| | Rion | Pre Amplifier | NH-25 | 76819 |
| | Rion | Microphone | UC-59 | 12814 |
| | Rion | Calibrator | NC-74 | 34536109 |
| | | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49
Procedures from IEC 61672-3:2006 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2002 YES Approval Number 21.21 / 13.02
If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 20 April 2018 ANV Job No. UKAS18/04261

Date Calibrated 20 April 2018

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

| Previous Certificate | Dated | Certificate No. | Laboratory |
|----------------------|---------------------|-----------------|------------|
| | Initial Calibration | | |

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number

UCRT18/1440

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | |
|--------------------------------------------------------------------------|-------------------|-----------------------------------------------|
| SLM instruction manual title | Sound Level Meter | NL-42 / NL-52 |
| SLM instruction manual ref / issue | | 11-03 |
| SLM instruction manual source | | Manufacturer |
| Internet download date if applicable | | N/A |
| Case corrections available | | Yes |
| Uncertainties of case corrections | | Yes |
| Source of case data | | Manufacturer |
| Wind screen corrections available | | Yes |
| Uncertainties of wind screen corrections | | Yes |
| Source of wind screen data | | Manufacturer |
| Mic pressure to free field corrections | | Yes |
| Uncertainties of Mic to F.F. corrections | | Yes |
| Source of Mic to F.F. corrections | | Manufacturer |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | | Yes |
| Specified or equivalent Calibrator | | Specified |
| Customer or Lab Calibrator | | Lab Calibrator |
| Calibrator adaptor type if applicable | | NC-74-002 |
| Calibrator cal. date | | 05 April 2018 |
| Calibrator cert. number | | UCRT18/1348 |
| Calibrator cal cert issued by | | 0653 |
| Calibrator SPL @ STP | 93.98 | dB Calibration reference sound pressure level |
| Calibrator frequency | 1001.90 | Hz Calibration check frequency |
| Reference level range | 25 - 130 | dB |

Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15
 Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 23.35 | 23.84 | ± 0.30 °C |
| Humidity | 38.4 | 35.6 | ± 3.00 %RH |
| Ambient Pressure | 101.37 | 101.36 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|------------------------------------------------------------------------------------|------|----|--------------------------|------|----|
| Initial indicated level | 94.0 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.
 Microphone installed (if requested by customer) = Less Than N/A dB A Weighting
 Uncertainty of the microphone installed self generated noise ± N/A dB

| Microphone replaced with electrical input device - | | UR = Under Range indicated | | | | | |
|----------------------------------------------------|------|----------------------------|------|-------|------|-------|--|
| Weighting | A | | C | | Z | | |
| | 11.6 | dB UR | 15.5 | dB UR | 21.4 | dB UR | |

Uncertainty of the electrical self generated noise ± 0.12 dB

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

..... END

Calibrated by: A Patel

R 1

Additional Comments

None



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 28 July 2017

Certificate Number: UCRT17/1627

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

| |
|--------------------|
| Page 1 of 2 Pages |
| Approved Signatory |
| |
| K. Mistry |

Customer Grenke Ltd
Unit 5D
Fifth Floor
Co. Cork
Ireland

Order No. 1741

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

| Identification | Manufacturer | Instrument | Type | Serial No. / Version |
|----------------|--------------|---------------------------------------|-------|----------------------|
| | Rion | Sound Level Meter | NL-52 | 00575785 |
| | Rion | Firmware | | 1.8 |
| | Rion | Pre Amplifier | NH-25 | 65812 |
| | Rion | Microphone | UC-59 | 11406 |
| | Rion | Calibrator | NC-74 | 34536109 |
| | | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2002 YES Approval Number 21.21 / 13.02

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 27 July 2017

ANV Job No. UKAS17/07375

Date Calibrated 28 July 2017

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

| Previous Certificate | Dated | Certificate No. | Laboratory |
|----------------------|---------------------|-----------------|------------|
| | Initial Calibration | | |

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number

UCRT17/1627

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | |
|--------------------------------------------------------------------------|-------------------|-----------------------------------------------|
| SLM instruction manual title | Sound Level Meter | NL-42 / NL-52 |
| SLM instruction manual ref / issue | | 11-03 |
| SLM instruction manual source | Manufacturer | |
| Internet download date if applicable | N/A | |
| Case corrections available | Yes | |
| Uncertainties of case corrections | Yes | |
| Source of case data | Manufacturer | |
| Wind screen corrections available | Yes | |
| Uncertainties of wind screen corrections | Yes | |
| Source of wind screen data | Manufacturer | |
| Mic pressure to free field corrections | Yes | |
| Uncertainties of Mic to F.F. corrections | Yes | |
| Source of Mic to F.F. corrections | Manufacturer | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | Yes | |
| Specified or equivalent Calibrator | Specified | |
| Customer or Lab Calibrator | Lab Calibrator | |
| Calibrator adaptor type if applicable | NC-74-002 | |
| Calibrator cal. date | 14 July 2017 | |
| Calibrator cert. number | UCRT17/1591 | |
| Calibrator cal cert issued by | 0653 | |
| Calibrator SPL @ STP | 94.03 | dB Calibration reference sound pressure level |
| Calibrator frequency | 1001.93 | Hz Calibration check frequency |
| Reference level range | 25 - 130 | dB |

Accessories used or corrected for during calibration - Wind Shield WS-10

Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|-------|-------|------------|
| Temperature | 22.56 | 22.97 | ± 0.20 °C |
| Humidity | 47.6 | 43.3 | ± 3.00 %RH |
| Ambient Pressure | 99.76 | 99.75 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|------------------------------------------------------------------------------------|------|----|--------------------------|------|----|
| Initial indicated level | 93.9 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|----------------------------------------------------------------|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

| Microphone replaced with electrical input device - | | UR = Under Range indicated | | | |
|------------------------------------------------------|------|----------------------------|------|----|----|
| Weighting | A | C | Z | | |
| | 13.5 | 15.6 | 21.1 | dB | UR |
| Uncertainty of the electrical self generated noise ± | | | 0.12 | dB | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

END

Calibrated by: A Patel

Additional Comments

None

R 1



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 28 July 2017

Certificate Number: UCRT17/1624

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

| |
|--------------------|
| Page 1 of 2 Pages |
| Approved Signatory |
| |
| K. Mistry |

Customer Grenke Ltd
 Unit 5D
 Fifth Floor
 Co. Cork
 Ireland

Order No. 1741

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

| Identification | Manufacturer | Instrument | Type | Serial No. / Version |
|----------------|--------------|---------------------------------------|-------|----------------------|
| | Rion | Sound Level Meter | NL-52 | 00575802 |
| | Rion | Firmware | | 1.8 |
| | Rion | Pre Amplifier | NH-25 | 75829 |
| | Rion | Microphone | UC-59 | 11426 |
| | Rion | Calibrator | NC-74 | 34536109 |
| | | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2002 YES Approval Number 21.21 / 13.02

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 27 July 2017

ANV Job No. UKAS17/07375

Date Calibrated 28 July 2017

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

| Previous Certificate | Dated | Certificate No. | Laboratory |
|----------------------|---------------------|-----------------|------------|
| | Initial Calibration | | |

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | |
|--------------------------------------------------------------------------|-------------------|-----------------------------------------------|
| SLM instruction manual title | Sound Level Meter | NL-42 / NL-52 |
| SLM instruction manual ref / issue | | 11-03 |
| SLM instruction manual source | Manufacturer | |
| Internet download date if applicable | N/A | |
| Case corrections available | Yes | |
| Uncertainties of case corrections | Yes | |
| Source of case data | Manufacturer | |
| Wind screen corrections available | Yes | |
| Uncertainties of wind screen corrections | Yes | |
| Source of wind screen data | Manufacturer | |
| Mic pressure to free field corrections | Yes | |
| Uncertainties of Mic to F.F. corrections | Yes | |
| Source of Mic to F.F. corrections | Manufacturer | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | Yes | |
| Specified or equivalent Calibrator | Specified | |
| Customer or Lab Calibrator | Lab Calibrator | |
| Calibrator adaptor type if applicable | NC-74-002 | |
| Calibrator cal. date | 14 July 2017 | |
| Calibrator cert. number | UCRT17/1591 | |
| Calibrator cal cert issued by | 0653 | |
| Calibrator SPL @ STP | 94.03 | dB Calibration reference sound pressure level |
| Calibrator frequency | 1001.93 | Hz Calibration check frequency |
| Reference level range | 25 - 130 | dB |

Accessories used or corrected for during calibration - Wind Shield WS-10
 Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|-------|-------|------------|
| Temperature | 21.72 | 22.23 | ± 0.20 °C |
| Humidity | 51.2 | 47.3 | ± 3.00 %RH |
| Ambient Pressure | 99.76 | 99.75 | ± 0.03 kPa |

| | | | |
|------------------------------------------------------------------------------------|------|----|--------------------------|
| Response to associated Calibrator at the environmental conditions above. | | | |
| Initial indicated level | 93.9 | dB | Adjusted indicated level |
| | | | 94.0 dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB |

| | | | |
|----------------------------------------------------------------|---------------------------------------------------|------------|-------------|
| Self Generated Noise | This test is currently not performed by this Lab. | | |
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |
| Microphone replaced with electrical input device - | UR = Under Range indicated | | |
| Weighting | A | C | Z |
| | 10.9 dB UR | 15.7 dB UR | 21.4 dB UR |
| Uncertainty of the electrical self generated noise ± | 0.12 dB | | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

..... END

Calibrated by: A Patel

Additional Comments

None

R 1