



# Appendix 10.7

## Calibration Certificates

Coolglass Wind Farm EIAR Volume 3

Coolglass Wind Farm Limited

SLR Project No.: 501.V00727.00006

26 June 2023



# CERTIFICATE OF CALIBRATION



0653


**Date of Issue: 05 October 2021**

**Certificate Number: UCRT21/2224**

Calibrated at & Certificate issued by:

ANV Measurement Systems  
 Beaufort Court  
 17 Roebuck Way  
 Milton Keynes MK5 8HL  
 Telephone 01908 642846 Fax 01908 642814  
 E-Mail: info@noise-and-vibration.co.uk  
 Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory  
B. Giles

Customer SLR Consulting Limited  
 2nd and 3rd Floors  
 15 Middle Pavement  
 Nottingham  
 NG1 7DX

Order No. 422-17278

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00710361
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	10903
	Rion	Microphone	UC-59	19635
	Rion	Calibrator	NC-75	34713324
		Calibrator adaptor type if applicable		NC-75-022

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

*Procedures from IEC 61672-3:2013 were used to perform the periodic tests.*

Type Approved to IEC 61672-1:2013 Yes

*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013*

Date Received 05 October 2021 ANV Job No. UKAS21/10653

Date Calibrated 05 October 2021

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	Initial Calibration		

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

<b>CERTIFICATE OF CALIBRATION</b>	<b>Certificate Number</b> UCRT21/2224
	Page 2 of 2 Pages

UKAS Accredited Calibration Laboratory No. 0653

**Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.**

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013	YES		
Specified or equivalent Calibrator	Specified		
Customer or Lab Calibrator	Customers Calibrator		
Calibrator adaptor type if applicable	NC-75-022		
Calibrator cal. date	05 October 2021		
Calibrator cert. number	UCRT21/2215		
Calibrator cal cert issued by Lab	0653		
Calibrator SPL @ STP	93.95	dB	Calibration reference sound pressure level
Calibrator frequency	1000.00	Hz	Calibration check frequency
Reference level range	Single dB		
Accessories used or corrected for during calibration -	Extension Cable & Wind Shield WS-15		
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests	Start	End	
Temperature	23.00	22.90	± 0.30 °C
Humidity	38.5	37.6	± 3.00 %RH
Ambient Pressure	99.25	99.33	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	93.9	dB	Adjusted indicated level
			94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	17.6	dB A Weighting
Microphone replaced with electrical input device -	UR = Under Range indicated		
Weighting	A	C	Z
	11.4 dB UR	15.2 dB UR	20.8 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

..... END .....

Calibrated by: C. Hirlav R 3



# CERTIFICATE OF CALIBRATION



0653

**Date of Issue: 05 October 2021**

**Certificate Number: UCRT21/2223**

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
B. Giles

Customer SLR Consulting Limited  
 2nd and 3rd Floors  
 15 Middle Pavement  
 Nottingham  
 NG1 7DX

Order No. 422-17278

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00710359
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	10901
	Rion	Microphone	UC-59	19633
	Rion	Calibrator	NC-75	34713324
		Calibrator adaptor type if applicable		NC-75-022

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

*Procedures from IEC 61672-3:2013 were used to perform the periodic tests.*

Type Approved to IEC 61672-1:2013 Yes

*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013*

Date Received 05 October 2021

ANV Job No. UKAS21/10653

Date Calibrated 05 October 2021

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	Initial Calibration		

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

<b>CERTIFICATE OF CALIBRATION</b>	<b>Certificate Number</b>
	<b>UCRT21/2223</b>
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

**Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.**

SLM instruction manual title NL-52/NL-42 Description for IEC 61672-1			
SLM instruction manual ref / issue No. 56034 21-03		Source Rion	
Date provided or internet download date 19 March 2021			
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator Specified			
Customer or Lab Calibrator		Customers Calibrator	
Calibrator adaptor type if applicable		NC-75-022	
Calibrator cal. date		05 October 2021	
Calibrator cert. number		UCRT21/2215	
Calibrator cal cert issued by Lab		0653	
Calibrator SPL @ STP	93.95	dB	Calibration reference sound pressure level
Calibrator frequency	1000.00	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests			
	Start	End	
Temperature	23.40	23.20	± 0.30 °C
Humidity	39.4	38.3	± 3.00 %RH
Ambient Pressure	99.21	99.24	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.0	dB	Adjusted indicated level 94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	17.0	dB A Weighting
Microphone replaced with electrical input device -		UR = Under Range indicated	
Weighting	A	C	Z
	11.8 dB UR	15.6 dB UR	22.8 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

**Additional Comments** The results on this certificate only relate to the items calibrated as identified above.

None

..... END .....

Calibrated by: B. Bogdan R 2



# CERTIFICATE OF CALIBRATION



0653

**Date of Issue: 05 October 2021**

**Certificate Number: UCRT21/2222**

Calibrated at & Certificate issued by:

ANV Measurement Systems  
 Beaufort Court  
 17 Roebuck Way  
 Milton Keynes MK5 8HL  
 Telephone 01908 642846 Fax 01908 642814  
 E-Mail: info@noise-and-vibration.co.uk  
 Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
B. Giles

Customer SLR Consulting Limited  
 2nd and 3rd Floors  
 15 Middle Pavement  
 Nottingham  
 NG1 7DX

Order No. 422-17278  
 Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator  
 Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00710358
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	10900
Rion	Microphone	UC-59	19632
Rion	Calibrator	NC-75	34713324
	Calibrator adaptor type if applicable		NC-75-022

Performance Class 1  
 Test Procedure TP 10. SLM 61672-3:2013  
*Procedures from IEC 61672-3:2013 were used to perform the periodic tests.*  
 Type Approved to IEC 61672-1:2013 Yes  
*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013*

Date Received 05 October 2021 ANV Job No. UKAS21/10653  
 Date Calibrated 05 October 2021

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
			Initial Calibration

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

<b>CERTIFICATE OF CALIBRATION</b>	<b>Certificate Number</b>
	<b>UCRT21/2222</b>
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

**Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.**

SLM instruction manual title NL-52/NL-42 Description for IEC 61672-1			
SLM instruction manual ref / issue No. 56034 21-03		Source Rion	
Date provided or internet download date 19 March 2021			
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013 YES			
Specified or equivalent Calibrator Specified			
Customer or Lab Calibrator		Customers Calibrator	
Calibrator adaptor type if applicable NC-75-022			
Calibrator cal. date 05 October 2021			
Calibrator cert. number UCRT21/2215			
Calibrator cal cert issued by Lab 0653			
Calibrator SPL @ STP		93.95 dB	Calibration reference sound pressure level
Calibrator frequency		1000.00 Hz	Calibration check frequency
Reference level range		Single	dB
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests			
	Start	End	
Temperature	22.80	23.20	± 0.30 °C
Humidity	40.4	39.5	± 3.00 %RH
Ambient Pressure	98.89	99.22	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.0 dB	Adjusted indicated level	94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±		0.10 dB	
Self Generated Noise			
Microphone installed -	Less Than	16.2 dB	A Weighting
Microphone replaced with electrical input device - UR = Under Range indicated			
Weighting	A	C	Z
	11.5 dB UR	14.6 dB UR	20.3 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

..... END .....

Calibrated by: C. Hirlav R 3



# CERTIFICATE OF CALIBRATION



0653


**Date of Issue: 05 October 2021**

**Certificate Number: UCRT21/2220**

Calibrated at & Certificate issued by:

ANV Measurement Systems  
 Beaufort Court  
 17 Roebuck Way  
 Milton Keynes MK5 8HL  
 Telephone 01908 642846 Fax 01908 642814  
 E-Mail: info@noise-and-vibration.co.uk  
 Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

B. Giles

Customer SLR Consulting Limited  
 2nd and 3rd Floors  
 15 Middle Pavement  
 Nottingham  
 NG1 7DX

Order No. 422-17278  
 Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator  
 Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00710362
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	10904
Rion	Microphone	UC-59	19636
Rion	Calibrator	NC-75	34713324
	Calibrator adaptor type if applicable		NC-75-022

Performance Class 1  
 Test Procedure TP 10. SLM 61672-3:2013  
*Procedures from IEC 61672-3:2013 were used to perform the periodic tests.*  
 Type Approved to IEC 61672-1:2013 Yes  
*If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013*  
 Date Received 05 October 2021 ANV Job No. UKAS21/10653  
 Date Calibrated 05 October 2021

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
			Initial Calibration

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



<b>CERTIFICATE OF CALIBRATION</b>	<b>Certificate Number</b>
	<b>UCRT21/2220</b>
UKAS Accredited Calibration Laboratory No. 0653	Page 2 of 2 Pages

**Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.**

SLM instruction manual title NL-52/NL-42 Description for IEC 61672-1			
SLM instruction manual ref / issue No. 56034 21-03		Source Rion	
Date provided or internet download date 19 March 2021			
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator Specified			
Customer or Lab Calibrator Customers Calibrator			
Calibrator adaptor type if applicable NC-75-022			
Calibrator cal. date 05 October 2021			
Calibrator cert. number UCRT21/2215			
Calibrator cal cert issued by Lab 0653			
Calibrator SPL @ STP 93.95 dB Calibration reference sound pressure level			
Calibrator frequency 1000.00 Hz Calibration check frequency			
Reference level range Single dB			
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests			
	Start	End	
Temperature	23.30	23.60	± 0.30 °C
Humidity	40.3	38.7	± 3.00 %RH
Ambient Pressure	99.02	99.11	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	93.9 dB	Adjusted indicated level	94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB

Self Generated Noise			
Microphone installed -	Less Than	17.1 dB	A Weighting
Microphone replaced with electrical input device - UR = Under Range indicated			
Weighting	A	C	Z
	12.2 dB UR	16.2 dB UR	23.4 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

**Additional Comments** The results on this certificate only relate to the items calibrated as identified above.

None

..... **END** .....

Calibrated by: **B. Bogdan** R 2



MTS Calibration Ltd,  
 The Grange Business Centre,  
 Belasis Avenue,  
 Billingham TS23 1LG,  
 England  
 Telephone: 01642 876 410

# CERTIFICATE OF CALIBRATION

Page 1 of 11 pages

Issued by: **MTS Calibration Ltd**

Approved Signatory:

Date of Issue: **12 August 2022** Certificate Number: **37295**

Tony Sherris

## Sound Level Meter

### Sound Level Meter Periodic Tests to EN 61672-3: 2013 Class 1

**Client:** Environmental Measurements  
 Unit 12, Tallaght Business Centre  
 Whitestown Business Park  
 Co.Dublin 24, Ireland

**Instrument Make:** Larson Davis  
**Instrument Model:** LxT1L  
**Serial Number:** 0004479

Associated Equipment	Make	Model	Serial number
Preamplifier	PCB	PRMLxT1L	042636
Microphone	PCB	377B02	339594
Calibrator	Brüel & Kjær	4231	3014620
Calibrator supplied by	MTS for this calibration		

The measurements were performed at The Grange Business Centre, Belasis Avenue, TS23 1LD. The results only apply to the items tested.

Periodic tests were performed in accordance with procedures from IEC 61672-3:2013 Class 1

Test results summary, detailed results are shown on subsequent pages.

Tests performed	Section	Results of test	Page	Comments
Calibration Certificate	22		1	
Additional Information			2	
Indication with Calibrator Supplied	10	No Limit	3	
Self-Generated Noise	11	No Limit	3	
Frequency and Time-weightings at 1kHz	14	Complies	3	
Long term stability	15	Complies	3	
High stability	21	Complies	3	
Acoustic Tests	12	Complies	4	
Frequency Weighting A	13	Complies	5	
Frequency Weighting C	13	Complies	6	
Frequency Weighting Z	13	Complies	7	
Level Linearity	16	Complies	8	
Level Linearity Range Control	17	n/a		SLM only has one range
Tone-burst Response	18	Complies	9	
Peak C sound level	19	Complies	10	
Overload indication	20	Complies	11	

The instrument required a replacement microphone in order to meet the above specifications.

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3: 2013 for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2: 2013, to demonstrate that the model of sound level meter fully conformed to the Class 1 specifications in IEC 61672-1: 2013, the sound level meter submitted for testing conforms to the Class 1 specifications of IEC 61672-1: 2013

Additional tests performed	Reference	
Microphone full frequency response	37310	See additional certificate
Filter calibration, third octave or octave	37295F	See additional certificate

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# Calibration Certificate

**Certificate Number 2021012148**

**Customer:**

Environmental Measurement  
Unit 12 Tallaght Business Centre  
Whitestown Business Park  
Dublin, 24, Ireland

<b>Model Number</b>	LxT SE	<b>Procedure Number</b>	D0001.8378
<b>Serial Number</b>	0006600	<b>Technician</b>	Ron Harris
<b>Test Results</b>	<b>Pass</b>	<b>Calibration Date</b>	28 Sep 2021
<b>Initial Condition</b>	As Manufactured	<b>Calibration Due</b>	
<b>Description</b>	Sound Expert LxT Class 1 Sound Level Meter Firmware Revision: 2.404	<b>Temperature</b>	23.76 °C ± 0.25 °C
		<b>Humidity</b>	50.2 %RH ± 2.0 %RH
		<b>Static Pressure</b>	85.46 kPa ± 0.13 kPa

**Evaluation Method** Tested electrically using Larson Davis PRMLxT1L S/N 070099 and a 12.0 pF capacitor to simulate microphone capacitance. Data reported in dB re 20 µPa assuming a microphone sensitivity of 23.6 mV/Pa.

**Compliance Standards** Compliant to Manufacturer Specifications and the following standards when combined with Calibration Certificate from procedure D0001.8384:

IEC 60651:2001 Type 1	ANSI S1.4-2014 Class 1
IEC 60804:2000 Type 1	ANSI S1.4 (R2006) Type 1
IEC 61252:2002	ANSI S1.25 (R2007)
IEC 61672:2013 Class 1	ANSI S1.43 (R2007) Type 1
IEC 61260:2001 Class 1	ANSI S1.11 (R2009) Class 1

Issuing lab certifies that the instrument described above meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST), or other national measurement institutes, and meets the requirements of ISO/IEC 17025:2017. **Test points marked with a ‡ in the uncertainties column do not fall within this laboratory's scope of accreditation.**

The quality system is registered to ISO 9001:2015.

This calibration is a direct comparison of the unit under test to the listed reference standards and did not involve any sampling plans to complete. No allowance has been made for the instability of the test device due to use, time, etc. Such allowances would be made by the customer as needed.

The uncertainties were computed in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM). A coverage factor of approximately 2 sigma (k=2) has been applied to the standard uncertainty to express the expanded uncertainty at approximately 95% confidence level.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the organization issuing this report.

Correction data from Larson Davis LxT Manual for SoundTrack LxT & SoundExpert Lxt, I770.01 Rev O Supporting Firmware Version 4.0.5, 2019-09-10

Calibration Check Frequency: 1000 Hz; Reference Sound Pressure Level: 114 dB re 20 µPa

LARSON DAVIS - A PCB PIEZOTRONICS DIV.  
1681 West 820 North  
Provo, UT 84601, United States  
716-684-0001



2021-9-28T15:22:40

Page 1 of 8

D0001.8407 Rev E



Making Sustainability Happen