

Issued to:

Fehily Timoney
J5 Plaza
North Park Business Park
North Road
Dublin 11

Calibration Reference

SLM200095

Test Date: 03/06/2020

Procedure: TP-SLM-1

Equipment

Item Calibrated:	Sound Level Meter	Model	977
Make:	Svantek	Serial Number:	69552

Calibration Procedure

The sound level meter was allowed to stabilize for a suitable period, as described in the manufacturer's instruction manual, in laboratory conditions. The sound level meter was calibrated by carrying out the verification tests detailed in IEC 61672-3 (2006), Periodic tests, specification of sound level meters. Tolerances for verification procedures are specified in IEC 61672-1 (2003).


Calibration Standards

Description	Serial Number
National Instruments PXI-4461	19C91D2
Stanford Research DS360	123803

The standards used in this calibration are traceable to NIST and/or other National Measurement Institutes (NMI's) that are signatories of the International Committee of Weights and Measures (CIPM) mutual recognition agreement (MRA).

Signed on behalf of Sonitus Systems:





Equipment Description

Model: Svantek **Serial Number:** 69552
Model: 977 **Microphone Model:** ACO 7052E

Ambient Conditions

Measurement conditions were within the tolerances defined in IEC 61672-1 and IEC 60942.

Barometric Pressure: 1030 hPa
Temperature: 21.6 °C
Relative Humidity: 45 %

Results Summary

IEC 61672 Test #	Test Description	Result
10	Self-generated noise	-
11	Frequency weighting (acoustical)	PASS
12	Frequency weighting (electrical)	PASS
13	Frequency and time weighting (1kHz)	PASS
14	Level linearity on reference level range	PASS
15	Level linearity with level range control	-
16	Toneburst response	PASS
17	Peak C sound level	PASS
18	Overload indication	PASS

As public evidence was available, from a testing organization responsible for approving the results of pattern evaluation tests, to demonstrate that the model of sound level meter fully conformed to the requirements for pattern evaluation described in IEC 61672:2003, the sound level meter tested is considered to conform to all the Class 1 requirements of IEC 61672:2003.

The manufacturer's guidelines concerning appropriate set up for measurement under various conditions should be observed during usage.

Prior to carrying out the verification tests the sound level meter was adjusted to read correctly using the acoustic calibrator held by the testing lab (Cirrus CR511ES, Serial number: 60871). The calibration procedure is described in the manufacturer's instruction manual.

Self-generated noise - IEC 61672-3 Test #10

SLM Measuring Mode: Leq

SLM Configuration	Freq. Weighting Network	SLM Reading
Microphone Installed	A	18.6
Microphone replaced by electrical input device fitted with short circuit	A	7.3
	C	7.3
	Z	12.6

Acoustical signal test of a frequency weighting - IEC 61672-3 Test #11

Range: reference level range

Frequency Weighting: C

Time Weighting: Slow

Input	Freq	Expected Level	Deviation	Tol +/-
94 dB	1000 Hz	94.0	0.0	1.0
	125 Hz	93.7	0.2	1.0
	4000 Hz	92.3	0.5	1.0

The frequency response was tested using an electrostatic actuator. Appropriate correction factors were applied where available from the manufacturer's instruction manual.

Electrical tests of frequency weighting - IEC 61672-3 Test #12

Range: reference level range

A-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.1	0.1	1.5	-1.5
125	95.0	95.0	0.0	1.5	-1.5
250	95.0	94.9	-0.1	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.1	0.1	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.2	0.2	2.1	-3.1
16000	95.0	94.8	-0.2	3.5	-17.0

C-weighting

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.0	0.0	1.5	-1.5
125	95.0	95.4	0.4	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.1	0.1	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.2	0.2	2.1	-3.1
16000	95.0	94.7	-0.3	3.5	-17.0

Linear

Freq	Expected Level	SLM Reading	Deviation	Tol +	Tol -
63	95.0	95.1	0.1	1.5	-1.5
125	95.0	95.1	0.1	1.5	-1.5
250	95.0	95.0	0.0	1.4	-1.4
500	95.0	95.0	0.0	1.4	-1.4
1000	95.0	95.0	0.0	1.1	-1.1
2000	95.0	95.0	0.0	1.6	-1.6
4000	95.0	95.1	0.1	1.6	-1.6
8000	95.0	95.1	0.1	2.1	-3.1
16000	95.0	95.1	0.1	3.5	-17.0

Frequency and Time Weightings at 1 kHz IEC 61672-3 Test #13

Range: reference level range

Time Weighting	Freq. Weighting	Expected Level	Deviation	Tol +/-
Fast	A	94.0	ref	
	C	94.0	0.0	0.2
Slow	A	94.0	0.0	0.2
	LEQ	A	94.0	0.0


Linearity level on reference range - IEC 61672-3 Test #14

Input frequency: 8 kHz

SLM Measuring Mode: SPL

Range	Expected Level	SLM Reading	Deviation	Tol +/-
123 dB	94.0	94.0	0.0	1.1
	99.0	99.0	0.0	1.1
	104.0	104.0	0.0	1.1
	109.0	109.0	0.0	1.1
	114.0	114.0	0.0	1.1
	119.0	119.0	0.0	1.1
	124.0	124.0	0.0	1.1
	129.0	129.0	0.0	1.1
	134.0	134.0	0.0	1.1
	135.0	135.0	0.0	1.1
	136.0	136.0	0.0	1.1
	137.0	137.0	0.0	1.1
	89.0	89.0	0.0	1.1
	84.0	84.0	0.0	1.1
	79.0	79.0	0.0	1.1
	74.0	74.0	0.0	1.1
	69.0	69.0	0.0	1.1
	64.0	64.0	0.0	1.1
	59.0	59.0	0.0	1.1
	54.0	54.0	0.0	1.1
	49.0	49.0	0.0	1.1
	44.0	44.0	0.0	1.1
	39.0	39.0	0.0	1.1
	38.0	38.0	0.0	1.1
	37.0	37.0	0.0	1.1
	36.0	36.0	0.0	1.1
	35.0	35.0	0.0	1.1

Offaly County Council, Planning Dept. Inspection Purposes Only



PL2 / 22 / 490
21 / 09 / 2022

Toneburst response - IEC 61672-3 Test #16

Range: reference level range

Burst Type	Response	Expected Level	SLM Reading	Deviation	Tol +	Tol -
0.25 ms	LAF _{MAX}	111.0	110.9	-0.1	0.8	-0.8
2.0 ms	LAF _{MAX}	120.0	119.9	-0.1	1.3	-1.3
200 ms	LAF _{MAX}	137.0	137.0	0.0	1.3	-3.3
2.0 ms	LAS _{MAX}	111.0	111.3	0.3	0.8	-0.8
200 ms	LAS _{MAX}	130.6	130.6	0.0	1.3	-3.3

Peak C sound level - IEC 61672-3 Test #17

Range: reference level range

Pulse Type	Freq	Expected Level	SLM Reading	Deviation	Tol +/-
1 cycle	8 kHz	135.4	135.3	-0.1	2.4
Pos ½ cycle	500 Hz	137.4	137.3	-0.1	1.4
Neg ½ cycle	500 Hz	137.4	137.3	-0.1	1.4

Overload indication IEC 61672-3 Test #18

Test Description	Overload at	Meas. Diff. (Pos – Neg)	Tol +/-
Pos. ½ cycle at 4 kHz	140.0		
Neg. ½ cycle at 4 kHz	140.0		
Level difference		0.0	1.8

Calibration Notes

1. The manufacturer's instruction manual was accessed through the manufacturer's website.
2. The sound level meter was powered by a regulated 9V power supply provided by the testing laboratory.