

RECEIVED: 29/11/2024



APPENDIX 10-1

CALIBRATION
RECORDS



CERTIFICATE OF CALIBRATION



0853

Date of Issue: 02 September 2022**Certificate Number: UCRT22/2053**

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court


17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.ukWeb: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Customer AWN Consulting Limited
The Tecpro Building
IDA Business and Technology Park
Clonshaugh
Dublin, D17 XD90
Ireland

Order No. 2243

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00976162
	Rion	Firmware		2.0
	Rion	Pre Amplifier	NH-25	76279
	Rion	Microphone	UC-59	12055
	Rion	Calibrator	NC-75	34313057
		Calibrator adaptor type if applicable		NC-75-022

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received 01 September 2022

ANV Job No. UKAS22/09555

Date Calibrated 02 September 2022

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	17 July 2020	UCRT20/1661	0853

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.