

RECEIVED: 02/09/2025



APPENDIX 12-2

**Copies of Calibration
Certificates**



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 21 March 2024

Certificate Number: UCRT24/1471

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way


Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Customer AWN Consulting
The Tecpro Building
IDA Business and Technology Park
Clonsaugh
Dublin 17
D17 XD90
Ireland

Order No. 2365

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00164426
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	54559
Rion	Microphone	UC-59	09207
Brüel & Kjær	Calibrator	4231	2022651
	Calibrator adaptor type if applicable		UC 0210

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013
Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes
If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received 20 March 2024 ANV Job No. UKAS24/03253

Date Calibrated 21 March 2024

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	04 November 2021	UCRT21/2362	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number

UCRT24/1471

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator	Equivalent		
Customer or Lab Calibrator	Customers Calibrator		
Calibrator adaptor type if applicable	UC 0210		
Calibrator cal. date	21 March 2024		
Calibrator cert. number	UCRT24/1467		
Calibrator cal cert issued by Lab	0653		
Calibrator SPL @ STP	94.01	dB	Calibration reference sound pressure level
Calibrator frequency	999.82	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			
Environmental conditions during tests	Start	End	
Temperature	23.16	22.76	± 0.30 °C
Humidity	56.4	55.9	± 3.00 %RH
Ambient Pressure	101.38	101.36	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.1	dB	Adjusted indicated level
			94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB
Self Generated Noise			
Microphone installed -	Less Than	18.8	dB A Weighting
Microphone replaced with electrical input device - UR = Under Range indicated			
Weighting	A	C	Z
	12.3 dB UR	15.6 dB UR	21.5 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

..... END
Calibrated by: K. Zablocki

R 1



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 13 June 2023

Certificate Number: UCRT23/1774

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way


Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Customer AWN Consulting Limited
The Tecpro Building
IDA Business and Technology Park
Clonshaugh
Dublin
D17 XD90
Ireland

Order No. AWN200423

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification	Manufacturer	Instrument	Type	Serial No. / Version
	Rion	Sound Level Meter	NL-52	00186668
	Rion	Firmware		2.1
	Rion	Pre Amplifier	NH-25	76701
	Rion	Microphone	UC-59	12813
	Rion	Calibrator	NC-74	34536109
		Calibrator adaptor type if applicable		NC-74-002

Performance Class 1

Test Procedure TP 10. SLM 61672-3:2013
Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013 Yes
If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received 13 June 2023 ANV Job No. UKAS23/06399

Date Calibrated 13 June 2023

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate	Dated	Certificate No.	Laboratory
	03 May 2022	UCRT22/1600	0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number

UCRT23/1774

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title NL-52/NL-42 Description for IEC 61672-1

SLM instruction manual ref / issue No. 56034 21-03 Source Rion

Date provided or internet download date 19 March 2021

Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
------------------	-------------------------	--

Uncertainties provided	Yes	Yes	Yes
------------------------	-----	-----	-----

Total expanded uncertainties within the requirements of IEC 61672-1:2013 YES

Specified or equivalent Calibrator Specified

Customer or Lab Calibrator Lab Calibrator

Calibrator adaptor type if applicable NC-74-002

Calibrator cal. date 30 May 2023

Calibrator cert. number UCRT23/1727

Calibrator cal cert issued by Lab 0653

Calibrator SPL @ STP 94.02 dB Calibration reference sound pressure level

Calibrator frequency 1001.99 Hz Calibration check frequency

Reference level range Single dB

Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15

Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.

Environmental conditions during tests	Start	End	
Temperature	24.08	24.03	± 0.30 °C
Humidity	41.0	35.8	± 3.00 %RH
Ambient Pressure	100.41	100.43	± 0.03 kPa

Indication at the Calibration Check Frequency

Initial indicated level	94.0 dB	Adjusted indicated level	94.0 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB

Self Generated Noise

Microphone installed - Less Than 19.2 dB A Weighting

Microphone replaced with electrical input device - UR = Under Range indicated

Weighting	A	C	Z
	14.5 dB UR	16.5 dB UR	22.0 dB UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

Prior to calibration the instrument's main PCB was replaced and the meter was realigned.

..... END
Calibrated by: K. Zablocki

R 1



CERTIFICATE OF CALIBRATION



0653

Date of Issue: 20 February 2024

Certificate Number: UCRT24/1283

Calibrated at & Certificate issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 2 Pages
Approved Signatory
K. Mistry

Customer AWN Consulting Limited
The Tecpro Building
IDA Business and Technology Park
Clonshaugh
Dublin
D17 XD90
Ireland

Order No.

2358

Description

Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification

Manufacturer	Instrument	Type	Serial No. / Version
Rion	Sound Level Meter	NL-52	00998409
Rion	Firmware		2.0
Rion	Pre Amplifier	NH-25	98623
Rion	Microphone	UC-59	15915
Brüel & Kjær	Calibrator	4231	2263026
	Calibrator adaptor type if applicable		UC 0210

Performance Class

1

Test Procedure

TP 10. SLM 61672-3:2013

Procedures from IEC 61672-3:2013 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2013

Yes

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013

Date Received

19 February 2024

ANV Job No.

UKAS24/02147

Date Calibrated

20 February 2024

The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.

Previous Certificate

Dated

01 February 2022

Certificate No.

UCRT22/1142

Laboratory

0653

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number

UCRT24/1283

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

SLM instruction manual title	NL-52/NL-42 Description for IEC 61672-1		
SLM instruction manual ref / issue	No. 56034 21-03	Source	Rion
Date provided or internet download date	19 March 2021		
	Case Corrections	Wind Shield Corrections	Mic Pressure to Free Field Corrections
Uncertainties provided	Yes	Yes	Yes
Total expanded uncertainties within the requirements of IEC 61672-1:2013			YES
Specified or equivalent Calibrator	Equivalent		
Customer or Lab Calibrator	Customers Calibrator		
Calibrator adaptor type if applicable	UC 0210		
Calibrator cal. date	20 February 2024		
Calibrator cert. number	UCRT24/1274		
Calibrator cal cert issued by Lab	0653		
Calibrator SPL @ STP	93.95	dB	Calibration reference sound pressure level
Calibrator frequency	999.97	Hz	Calibration check frequency
Reference level range	Single	dB	
Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15			
Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.			

Environmental conditions during tests	Start	End	
Temperature	22.58	22.72	± 0.30 °C
Humidity	48.3	53.9	± 3.00 %RH
Ambient Pressure	101.11	101.10	± 0.03 kPa

Indication at the Calibration Check Frequency			
Initial indicated level	94.2	dB	Adjusted indicated level 93.9 dB
Uncertainty of calibrator used for Indication at the Calibration Check Frequency ±			0.10 dB

Self Generated Noise			
Microphone installed -	Less Than	18.1	dB A Weighting
Microphone replaced with electrical input device -		UR = Under Range indicated	
Weighting	A	C	Z
	11.9	15.0	20.4
	dB	dB	dB
	UR	UR	UR

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

END

Calibrated by: K. Zablocki

R 1